

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020321\
 Data File : VL036297.D
 Acq On : 3 Feb 2021 17:19
 Operator : SY/AP
 Sample : M1295-01DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-INFLUENTDL

Manual Integrations
 APPROVED

MMDadoda
 2/4/2021 5:15:59 PM

Quant Time: Feb 04 01:47:09 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.690	49	1597046	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.201	114	3448886	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.118	117	3164463	10.00	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.452	95	2630817	10.04	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%
Target Compounds						
						Qvalue
13) Ethanol	3.636	45	50495m	3.44	ppbv	
31) cis-1,2-Dichloroethene	5.565	61	23576m	0.15	ppbv	
38) Trichloroethene	7.861	130	18895m	0.16	ppbv	
52) Tetrachloroethene	11.259	164	257137	2.21	ppbv	94
53) Toluene	9.835	91	74327	0.21	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020321\
Data File : VL036297.D
Acq On : 3 Feb 2021 17:19
Operator : SY/AP
Sample : M1295-01DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-INFLUENTDL

Manual Integrations
APPROVED

MMDadoda
2/4/2021 5:15:59 PM

Quant Time: Feb 04 01:47:09 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

