

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020321\
 Data File : VL036298.D
 Acq On : 3 Feb 2021 17:57
 Operator : SY/AP
 Sample : M1295-06 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OFFSITE-DDC-EFFLUENT

Manual Integrations
 APPROVED

MMDadoda
 2/4/2021 5:15:57 PM

Quant Time: Feb 04 01:47:20 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.693	49	1605321	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.201	114	3408282	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.118	117	3124447	10.00	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.449	95	2495752	9.65	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.50%
Target Compounds						
						Qvalue
9) Propene	3.038	41	12945m	0.14	ppbv	
13) Ethanol	3.636	45	25261m	1.71	ppbv	
15) Acetone	3.886	43	84250m	0.43	ppbv	
20) Methylene Chloride	4.366	84	58988	0.94	ppbv	99
26) Hexane	5.710	57	31741m	0.19	ppbv	
31) cis-1,2-Dichloroethene	5.574	61	127992	0.81	ppbv	94
34) 2-Butanone	5.285	43	35096m	0.14	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020321\
Data File : VL036298.D
Acq On : 3 Feb 2021 17:57
Operator : SY/AP
Sample : M1295-06 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OFFSITE-DDC-EFFLUENT

Manual Integrations
APPROVED

MMDadoda
2/4/2021 5:15:57 PM

Quant Time: Feb 04 01:47:20 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

