

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL020419\
 Data File : VL033162.D
 Acq On : 4 Feb 2019 19:53
 Operator : SY/AP
 Sample : K1267-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSA2DL

Manual Integrations
 APPROVED

MMDadoda
 2/5/2019 10:28:09 AM

Quant Time: Feb 05 01:56:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL020419AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Feb 05 2019
 QLast Update : Mon Feb 04 14:41:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1507062	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3498887	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	3365219	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2508647	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	10856m	0.129	ppbv	
10) Heptane	8.25	43	38711	0.147	ppbv	96
13) Ethanol	3.67	45	20324m	0.611	ppbv	
15) Acetone	3.90	43	1380560	7.694	ppbv	99
26) Hexane	5.78	57	23066m	0.133	ppbv	
53) Toluene	9.95	91	719475	1.713	ppbv	100
59) m/p-Xylene	13.12	91	130343m	0.301	ppbv	
60) o-Xylene	13.86	91	44269m	0.111	ppbv	
62) Isopropylbenzene	14.83	105	459725	0.805	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL020419\
Data File : VL033162.D
Acq On : 4 Feb 2019 19:53
Operator : SY/AP
Sample : K1267-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampled :
SSA2DL

Manual Integrations
APPROVED

MMDadoda
2/5/2019 10:28:09 AM

Quant Time: Feb 05 01:56:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL020419AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Feb
QLast Update : Mon Feb 04 14:41:39 2019
Response via : Initial Calibration

