

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020924\
 Data File : VL041064.D
 Acq On : 9 Feb 2024 19:04
 Operator : SY/MD
 Sample : P1420-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2RDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/12/2024
 Supervised By :Mahesh Dadoda 02/12/2024

Quant Time: Feb 12 00:49:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Feb 12 00:40:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1228351	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.648	114	3246730	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2820933	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	2142095	10.054	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.500%
Target Compounds						
						Qvalue
13) Ethanol	3.237	45	625301	38.716	ppbv	93
15) Acetone	3.475	43	114857	0.731	ppbv	100
19) Isopropyl Alcohol	3.587	45	107493m	1.182	ppbv	
20) Methylene Chloride	3.931	84	173598	2.872	ppbv	91
26) Hexane	5.211	57	142358	1.173	ppbv #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020924\
Data File : VL041064.D
Acq On : 9 Feb 2024 19:04
Operator : SY/MD
Sample : P1420-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2RDL

Quant Time: Feb 12 00:49:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Feb 12 00:40:04 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/12/2024
Supervised By :Mahesh Dadoda 02/12/2024

