

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL021319\
 Data File : VL033186.D
 Acq On : 13 Feb 2019 13:53
 Operator : SY/AP
 Sample : VL0213ABL01
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0213ABL01

Manual Integrations
 APPROVED

MMDadoda
 2/20/2019 12:23:26 PM

Quant Time: Feb 20 07:16:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 20 07:16:01 2019
 QLast Update : Wed Feb 20 06:00:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	977217	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2246978	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	2283006	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2048125	11.51	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	115.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
15) Acetone	3.93	43	12936m	0.117	ppbv	
20) Methylene Chloride	4.41	84	5388m	0.123	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL021319\
Data File : VL033186.D
Acq On : 13 Feb 2019 13:53
Operator : SY/AP
Sample : VL0213ABL01
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0213ABL01

Manual Integrations
APPROVED

MMDadoda
2/20/2019 12:23:26 PM

Quant Time: Feb 20 07:16:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL020519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 20 06:00:52 2019
Response via : Initial Calibration

