

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021319\
 Data File : VL033198.D
 Acq On : 13 Feb 2019 22:20
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 14 06:13:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 06 09:13:11 2019
 QLast Update : Wed Feb 06 09:13:11 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	1.912	1.492	22.0	78	0.00
3	Chlorodifluoromethane	1.204	1.022	15.1	86	0.00
4	Chloromethane	0.636	0.664	-4.4	103	0.00
5 T	Vinyl Chloride	0.664	0.644	3.0	105	0.00
6 T	Bromomethane	0.365	0.378	-3.6	102	0.00
7	Chloroethane	0.217	0.232	-6.9	106	0.00
8 T	Dichlorotetrafluoroethane	1.495	1.533	-2.5	107	0.00
9 T	Propene	0.582	0.467	19.8	83	0.00
10 T	Heptane	1.760	1.420	19.3	84	0.00
11 T	Trichlorofluoromethane	1.548	1.667	-7.7	110	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.073	1.065	0.7	101	0.00
13	Ethanol	0.194	0.134	30.9#	97	0.00
14 T	Bromoethene	0.463	0.479	-3.5	103	0.00
15 T	Acetone	1.127	1.071	5.0	104	0.00
16 T	1,3-Butadiene	0.535	0.541	-1.1	102	0.00
17	tert-Butyl alcohol	0.215	0.211	1.9	95	0.00
18 T	1,1-Dichloroethene	0.480	0.476	0.8	97	0.00
19 T	Isopropyl Alcohol	0.764	0.745	2.5	107	0.00
20 T	Methylene Chloride	0.448	0.411	8.3	99	0.00
21 T	Allyl Chloride	0.744	0.760	-2.2	97	0.00
22 T	trans-1,2-Dichloroethene	0.466	0.470	-0.9	99	0.00
23 T	Vinyl Acetate	1.953	1.957	-0.2	99	0.00
24 T	1,1-Dichloroethane	1.381	1.343	2.8	101	0.00
25 T	Ethyl Acetate	2.533	2.368	6.5	103	0.00
26 T	Hexane	1.189	1.049	11.8	101	0.00
27 T	Carbon Disulfide	1.089	1.168	-7.3	97	0.00
28 T	Methyl tert-Butyl Ether	1.625	1.540	5.2	97	0.00
29 T	Chloroform	1.822	1.707	6.3	100	0.00
30 T	Cyclohexane	1.103	0.905	18.0	84	0.00
31 T	cis-1,2-Dichloroethene	1.190	1.119	6.0	100	0.00
32 T	1,1,1-Trichloroethane	2.007	1.650	17.8	100	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
34 T	2-Butanone	0.634	0.700	-10.4	101	0.00
35 T	Carbon Tetrachloride	0.764	0.750	1.8	87	0.00
36 T	Benzene	1.013	0.954	5.8	92	0.00
37 T	1,2-Dichloroethane	0.562	0.595	-5.9	106	0.00
38 T	Trichloroethene	0.434	0.355	18.2	86	0.00
39 T	1,2-Dichloropropane	0.400	0.368	8.0	82	0.00
40 T	1,4-Dioxane	0.145	0.127	12.4	84	0.00
41 T	Tetrahydrofuran	0.393	0.400	-1.8	97	0.00
42 T	Bromodichloromethane	0.814	0.822	-1.0	88	0.00
43	Methyl Methacrylate	0.407	0.401	1.5	85	0.00
44 T	2,2,4-Trimethylpentane	1.717	1.592	7.3	86	0.00
45 T	t-1,3-Dichloropropene	0.452	0.527	-16.6	85	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021319\
 Data File : VL033198.D
 Acq On : 13 Feb 2019 22:20
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 14 06:13:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 06 09:13:11 2019
 QLast Update : Wed Feb 06 09:13:11 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.576	0.607	-5.4	85	0.00
47 T	1,1,2-Trichloroethane	0.415	0.395	4.8	85	0.00
48 T	Dibromochloromethane	0.654	0.691	-5.7	87	0.00
49 T	Bromoform	0.563	0.619	-9.9	89	0.00
50 T	4-Methyl-2-Pentanone	1.017	1.017	0.0	88	0.00
51 T	2-Hexanone	0.802	0.794	1.0	84	0.00
52 T	Tetrachloroethene	0.389	0.352	9.5	87	0.00
53 T	Toluene	1.200	1.144	4.7	84	0.00
54 T	1,2-Dibromoethane	0.651	0.633	2.8	86	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
56	1,1,1,2-Tetrachloroethane	0.459	0.460	-0.2	88	0.00
57 T	Chlorobenzene	0.873	0.796	8.8	85	0.00
58 T	Ethyl Benzene	1.618	1.548	4.3	87	0.00
59 T	m/p-Xylene	1.351	1.274	5.7	87	0.00
60 T	o-Xylene	1.313	1.225	6.7	88	0.00
61 T	Styrene	0.991	0.936	5.5	84	0.00
62	Isopropylbenzene	1.853	1.711	7.7	85	0.00
63 T	1,1,2,2-Tetrachloroethane	1.032	0.838	18.8	87	0.00
64	n-propylbenzene	0.471	0.436	7.4	84	0.00
65	tert-Butylbenzene	1.613	1.482	8.1	85	0.00
66 T	Benzyl Chloride	0.573	0.781	-36.3#	86	0.00
67	sec-Butylbenzene	2.342	2.182	6.8	87	0.00
68 S	1-Bromo-4-Fluorobenzene	0.770	0.782	-1.6	79	0.00
69	p-Isopropyltoluene	1.951	1.767	9.4	85	0.00
70	n-Butylbenzene	2.105	1.929	8.4	87	0.00
71	2-Chlorotoluene	1.456	1.340	8.0	85	0.00
72 T	4-Ethyltoluene	1.514	1.417	6.4	84	0.00
73 T	1,3,5-Trimethylbenzene	1.448	1.365	5.7	85	0.00
74 T	1,2,4-Trimethylbenzene	1.503	1.386	7.8	87	0.00
75 T	1,3-Dichlorobenzene	0.867	0.804	7.3	86	0.00
76 T	1,4-Dichlorobenzene	0.858	0.804	6.3	85	0.00
77 T	1,2-Dichlorobenzene	0.856	0.779	9.0	84	0.00
78 T	Hexachloro-1,3-Butadiene	0.426	0.631	-48.1#	149	0.00
79 T	Naphthalene	1.614	1.519	5.9	87	0.00
80 T	Naphthalene,2-methyl-	0.817	0.717	12.2	88	0.00
81 T	1,2,4-Trichlorobenzene	0.820	0.740	9.8	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0