

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021920\
 Data File : VL034649.D
 Acq On : 19 Feb 2020 12:52
 Operator : SY/AP
 Sample : VL0219ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0219ABL01

Manual Integrations
 APPROVED

MMDadoda
 2/21/2020 3:23:59 PM

Quant Time: Feb 20 07:13:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 30 2020
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Internal Standards			R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane			5.65	49	766147	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene			7.17	114	1589284	10.00	ppbv	-0.01
55) Chlorobenzene-d5			12.11	117	1626290	10.00	ppbv	0.00
System Monitoring Compounds								
68) 1-Bromo-4-Fluorobenzene			14.45	95	1381149	10.05	ppbv	0.00
Spiked Amount		10.000	Range	65 - 135		Recovery	=	100.50%
Target Compounds								
15) Acetone			3.84	43	41402	0.368	ppbv	Ovalue # 75
20) Methylene Chloride			4.32	84	15592m	0.364	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021920\
Data File : VL034649.D
Acq On : 19 Feb 2020 12:52
Operator : SY/AP
Sample : VL0219ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0219ABL01

Manual Integrations
APPROVED

MMDadoda
2/21/2020 3:23:59 PM

Quant Time: Feb 20 07:13:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 30 17:16:44 2020
Response via : Initial Calibration

