

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021920\
 Data File : VL034664.D
 Acq On : 20 Feb 2020 9:10
 Operator : SY/AP
 Sample : QC021920 CAN 10296
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC021920 CAN 10296

Quant Time: Feb 20 14:52:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Jan 30 17:16:44 2020
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	895048	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.17	114	1862050	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.10	117	1857712	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1524016	9.71	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021920\
Data File : VL034664.D
Acq On : 20 Feb 2020 9:10
Operator : SY/AP
Sample : QC021920 CAN 10296
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC021920 CAN 10296

Quant Time: Feb 20 14:52:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 30 17:16:44 2020
QLast Update : Thu Jan 30 17:16:44 2020
Response via : Initial Calibration

