

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL021920\
 Data File : VL034648.D
 Acq On : 19 Feb 2020 11:08
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 20 07:07:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 30 17:16:44 2020
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	2.203	1.562	29.1	65	0.00
3	Chlorodifluoromethane	1.208	1.241	-2.7	90	0.00
4	Chloromethane	0.739	0.749	-1.4	90	0.00
5 T	Vinyl Chloride	0.706	0.723	-2.4	91	0.00
6 T	Bromomethane	0.367	0.397	-8.2	93	0.00
7	Chloroethane	0.263	0.275	-4.6	91	0.00
8 T	Dichlorotetrafluoroethane	1.722	1.768	-2.7	94	0.00
9 T	Propene	0.470	0.462	1.7	85	0.00
10 T	Heptane	1.184	1.277	-7.9	85	0.00
11 T	Trichlorofluoromethane	2.001	2.094	-4.6	93	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.305	1.409	-8.0	95	0.00
13	Ethanol	0.126	0.127	-0.8	101	0.00
14 T	Bromoethene	0.512	0.564	-10.2	93	0.00
15 T	Acetone	1.470	1.464	0.4	94	0.00
16 T	1,3-Butadiene	0.584	0.650	-11.3	91	0.00
17	tert-Butyl alcohol	1.081	1.403	-29.8	106	0.00
18 T	1,1-Dichloroethene	0.551	0.625	-13.4	96	0.00
19 T	Isopropyl Alcohol	0.858	1.064	-24.0	106	0.00
20 T	Methylene Chloride	0.560	0.545	2.7	94	0.00
21 T	Allyl Chloride	0.913	0.998	-9.3	91	0.00
22 T	trans-1,2-Dichloroethene	0.577	0.630	-9.2	93	0.00
23 T	Vinyl Acetate	1.677	1.769	-5.5	86	0.00
24 T	1,1-Dichloroethane	1.400	1.461	-4.4	90	0.00
25 T	Ethyl Acetate	2.379	2.432	-2.2	88	0.00
26 T	Hexane	0.997	1.026	-2.9	88	0.00
27 T	Carbon Disulfide	1.345	1.597	-18.7	93	0.00
28 T	Methyl tert-Butyl Ether	1.700	1.820	-7.1	89	0.00
29 T	Chloroform	1.762	1.806	-2.5	90	0.00
30 T	Cyclohexane	0.739	0.813	-10.0	88	0.00
31 T	cis-1,2-Dichloroethene	1.035	1.117	-7.9	88	0.00
32 T	1,1,1-Trichloroethane	1.892	1.895	-0.2	91	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
34 T	2-Butanone	0.726	0.734	-1.1	86	0.00
35 T	Carbon Tetrachloride	0.956	1.018	-6.5	91	0.00
36 T	Benzene	0.889	0.962	-8.2	88	0.00
37 T	1,2-Dichloroethane	0.742	0.791	-6.6	92	0.00
38 T	Trichloroethene	0.375	0.423	-12.8	89	0.00
39 T	1,2-Dichloropropane	0.369	0.392	-6.2	89	0.00
40 T	1,4-Dioxane	0.142	0.157	-10.6	92	0.00
41 T	Tetrahydrofuran	0.379	0.397	-4.7	85	0.00
42 T	Bromodichloromethane	0.913	0.981	-7.4	89	0.00
43	Methyl Methacrylate	0.358	0.399	-11.5	87	0.00
44 T	2,2,4-Trimethylpentane	1.613	1.711	-6.1	88	0.00
45 T	t-1,3-Dichloropropene	0.468	0.579	-23.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL021920\
 Data File : VL034648.D
 Acq On : 19 Feb 2020 11:08
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 20 07:07:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 30 17:16:44 2020
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.532	0.644	-21.1	88	0.00
47 T	1,1,2-Trichloroethane	0.399	0.424	-6.3	89	0.00
48 T	Dibromochloromethane	0.732	0.839	-14.6	89	0.00
49 T	Bromoform	0.642	0.739	-15.1	88	0.00
50 T	4-Methyl-2-Pentanone	1.035	1.125	-8.7	88	0.00
51 T	2-Hexanone	0.815	0.923	-13.3	87	0.00
52 T	Tetrachloroethene	0.357	0.385	-7.8	91	0.00
53 T	Toluene	0.977	1.139	-16.6	90	0.00
54 T	1,2-Dibromoethane	0.598	0.654	-9.4	90	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
56	1,1,1,2-Tetrachloroethane	0.494	0.538	-8.9	90	0.00
57 T	Chlorobenzene	0.823	0.890	-8.1	91	0.00
58 T	Ethyl Benzene	1.338	1.517	-13.4	89	0.00
59 T	m/p-Xylene	1.183	1.284	-8.5	89	0.00
60 T	o-Xylene	1.188	1.303	-9.7	89	0.00
61 T	Styrene	0.792	0.941	-18.8	88	0.00
62	Isopropylbenzene	1.539	1.677	-9.0	89	0.00
63 T	1,1,2,2-Tetrachloroethane	0.919	0.906	1.4	88	0.00
64	n-propylbenzene	0.396	0.435	-9.8	87	0.00
65	tert-Butylbenzene	1.395	1.534	-10.0	91	0.00
66 T	Benzyl Chloride	0.585	0.741	-26.7	88	0.00
67	sec-Butylbenzene	1.917	2.086	-8.8	91	0.00
68 S	1-Bromo-4-Fluorobenzene	0.845	0.862	-2.0	79	0.00
69	p-Isopropyltoluene	1.556	1.742	-12.0	90	0.00
70	n-Butylbenzene	1.652	1.817	-10.0	91	0.00
71	2-Chlorotoluene	1.169	1.288	-10.2	87	0.00
72 T	4-Ethyltoluene	1.305	1.475	-13.0	90	0.00
73 T	1,3,5-Trimethylbenzene	1.275	1.418	-11.2	90	0.00
74 T	1,2,4-Trimethylbenzene	1.412	1.501	-6.3	91	0.00
75 T	1,3-Dichlorobenzene	0.813	0.839	-3.2	90	0.00
76 T	1,4-Dichlorobenzene	0.801	0.831	-3.7	90	0.00
77 T	1,2-Dichlorobenzene	0.781	0.795	-1.8	90	0.00
78 T	Hexachloro-1,3-Butadiene	0.747	0.749	-0.3	92	0.00
79 T	Naphthalene	1.111	1.267	-14.0	91	0.00
80 T	Naphthalene,2-methyl-	0.428	0.675	-57.7#	103	0.00
81 T	1,2,4-Trichlorobenzene	0.685	0.776	-13.3	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0