

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022322\
 Data File : VL038406.D
 Acq On : 23 Feb 2022 12:44
 Operator : SY/AP
 Sample : VL0223ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0223ABL01

Quant Time: Feb 24 07:26:13 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL02222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	972286	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.179	114	2280176	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.085	117	1938345	10.000	ppbv	# 0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.413	95	1576697	9.831	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022322\
Data File : VL038406.D
Acq On : 23 Feb 2022 12:44
Operator : SY/AP
Sample : VL0223ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0223ABL01

Quant Time: Feb 24 07:26:13 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

