

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022323\
 Data File : VL040226.D
 Acq On : 24 Feb 2023 9:16
 Operator : SY/MD
 Sample : QC022123 CAN 10605
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC022123 CAN 10605

Quant Time: Feb 24 10:17:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Feb 24 00:45:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	699131	10.000	ppbv	-0.03
33) 1,4-Difluorobenzene	6.635	114	1768255	10.000	ppbv	-0.04
55) Chlorobenzene-d5	11.455	117	1585749	10.000	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.764	95	1304433	10.004	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022323\
Data File : VL040226.D
Acq On : 24 Feb 2023 9:16
Operator : SY/MD
Sample : QC022123 CAN 10605
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC022123 CAN 10605

Quant Time: Feb 24 10:17:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Feb 24 00:45:34 2023
Response via : Initial Calibration

