

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022323\
 Data File : VL040217.D
 Acq On : 23 Feb 2023 11:10
 Operator : SY/MD
 Sample : VL0223ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0223ABL01

Quant Time: Feb 24 00:46:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Feb 24 00:45:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	830784	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2209058	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	1938830	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1569123	9.842	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.400%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022323\
Data File : VL040217.D
Acq On : 23 Feb 2023 11:10
Operator : SY/MD
Sample : VL0223ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0223ABL01

Quant Time: Feb 24 00:46:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Feb 24 00:45:34 2023
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.192 min Scan# 801
Delta R.T. 0.003 min
Lab File: VL040217.D
Acq: 23 Feb 2023 11:10




Instrument :
MSVOA_L
ClientSampleId :
VL0223ABL01


Tgt Ion: 49 Resp: 830784
Ion Ratio Lower Upper
49 100
128 51.0 24.1 72.3
130 63.3 30.9 92.8



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 6.655 min Scan# 1260
Delta R.T. 0.000 min
Lab File: VL040217.D
Acq: 23 Feb 2023 11:10

Tgt Ion: 114 Resp: 2209058
Ion Ratio Lower Upper
114 100
63 25.6 20.4 30.6
88 17.9 14.3 21.5





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 11.471 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL040217.D

Acq: 23 Feb 2023 11:10

Instrument :

MSVOA_1

ClientSampleId :

VL0223ABL01





Tgt Ion:117 Resp: 1938830

Ion Ratio Lower Upper

117 100

82 68.5 53.3 79.9

119 33.0 24.6 37.0



#68

1-Bromo-4-Fluorobenzene

Concen: 9.842 ppbv

RT: 13.783 min Scan# 3477

Delta R.T. 0.000 min

Lab File: VL040217.D

Acq: 23 Feb 2023 11:10

Tgt Ion: 95 Resp: 1569123

Ion Ratio Lower Upper

95 100

174 59.3 48.2 72.2

175 4.2 3.4 5.2

