

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022521\
 Data File : VL036386.D
 Acq On : 25 Feb 2021 18:16
 Operator : SY/AP
 Sample : M1463-14DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NS-6DL

Manual Integrations
 APPROVED

MMDadoda
 3/1/2021 10:02:52 AM

Quant Time: Feb 26 15:44:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1564734	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3320965	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	2959789	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2419034	9.93	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	31188m	2.291	ppbv	
15) Acetone	3.85	43	85312	0.446	ppbv #	60
20) Methylene Chloride	4.33	84	187291	3.035	ppbv	93
26) Hexane	5.67	57	248876	1.640	ppbv #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022521\
Data File : VL036386.D
Acq On : 25 Feb 2021 18:16
Operator : SY/AP
Sample : M1463-14DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NS-6DL

Manual Integrations
APPROVED

MMDadoda
3/1/2021 10:02:52 AM

Quant Time: Feb 26 15:44:37 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

