


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022521\
 Data File : VL036376.D
 Acq On : 25 Feb 2021 9:23
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL021021AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Feb 10 16:04:02 2021
 Last Update : Wed Feb 10 16:04:02 2021



AutoFind: Scans 3613, 3614, 3615; Background Corrected with Scan 3591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	26.2	240533	PASS
75	95	30	66	60.0	550229	PASS
95	95	100	100	100.0	917418	PASS
96	95	5	9	7.2	66077	PASS
173	174	0.00	2	1.1	5950	PASS
174	95	50	120	60.4	554090	PASS
175	174	4	9	7.7	42461	PASS
176	174	93	101	96.5	534464	PASS
177	176	5	9	6.3	33701	PASS