

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022621\
 Data File : VL036401.D
 Acq On : 26 Feb 2021 14:10
 Operator : SY/AP
 Sample : M1502-03 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-6

Manual Integrations
 APPROVED

MMDadoda
 3/1/2021 10:08:27 AM

Quant Time: Feb 27 03:37:05 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10 2021
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1480740	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.14	114	3189696	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.05	117	2917159	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	2262178	9.43	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	25444m	0.371	ppbv	
13) Ethanol	3.60	45	39003m	3.027	ppbv	
15) Acetone	3.83	43	184927	1.022	ppbv	# 73
20) Methylene Chloride	4.32	84	1257637	21.537	ppbv	98
26) Hexane	5.66	57	140261m	0.977	ppbv	
34) 2-Butanone	5.23	43	26561m	0.125	ppbv	
53) Toluene	9.77	91	64538m	0.220	ppbv	
59) m/p-Xylene	12.93	91	64222m	0.183	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022621\
Data File : VL036401.D
Acq On : 26 Feb 2021 14:10
Operator : SY/AP
Sample : M1502-03 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-6

Manual Integrations
APPROVED

MMDadoda
3/1/2021 10:08:27 AM

Quant Time: Feb 27 03:37:05 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

