

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022720\
 Data File : VL034708.D
 Acq On : 27 Feb 2020 14:33
 Operator : SY/AP
 Sample : L1662-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IAS-1-M-ELECT-RMDL

Manual Integrations
 APPROVED

MMDadoda
 2/28/2020 12:32:34 PM

Quant Time: Feb 28 01:30:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 26 06:58:15 2020
 QLast Update : Wed Feb 26 06:58:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1021558	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.20	114	2072832	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.14	117	2062572	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1710930	10.12	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	8268m	0.179	ppbv	
13) Ethanol	3.60	45	18206	1.251	ppbv	99
15) Acetone	3.85	43	51638	0.444	ppbv	93
20) Methylene Chloride	4.34	84	23150	0.535	ppbv #	87
26) Hexane	5.69	57	15039m	0.158	ppbv	
53) Toluene	9.85	91	48629m	0.255	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022720\
Data File : VL034708.D
Acq On : 27 Feb 2020 14:33
Operator : SY/AP
Sample : L1662-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IAS-1-M-ELECT-RMDL

Manual Integrations
APPROVED

MMDadoda
2/28/2020 12:32:34 PM

Quant Time: Feb 28 01:30:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 26 06:58:15 2020
Response via : Initial Calibration

