

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022720\
 Data File : VL034727.D
 Acq On : 28 Feb 2020 3:14
 Operator : SY/AP
 Sample : L1661-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-4-STORAGE-ROOM-NEXT-TO-C

Manual Integrations
 APPROVED

MMDadoda
 2/28/2020 12:33:44 PM

Quant Time: Feb 28 03:51:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1143921	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2374878	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2339932	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1926556	10.04	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	194283	11.918	ppbv	97
15) Acetone	3.84	43	166679	1.279	ppbv	94
17) tert-Butyl alcohol	4.29	59	33440m	0.335	ppbv	
19) Isopropyl Alcohol	3.96	45	297647	3.834	ppbv	100
20) Methylene Chloride	4.33	84	23130	0.477	ppbv #	82
26) Hexane	5.68	57	13131m	0.123	ppbv	
34) 2-Butanone	5.25	43	25937m	0.174	ppbv	
38) Trichloroethene	7.85	130	46392	0.587	ppbv	98
41) Tetrahydrofuran	6.07	42	12863m	0.163	ppbv	
53) Toluene	9.83	91	47576m	0.218	ppbv	
59) m/p-Xylene	13.01	91	26305m	0.107	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022720\
Data File : VL034727.D
Acq On : 28 Feb 2020 3:14
Operator : SY/AP
Sample : L1661-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-4-STORAGE-ROOM-NEXT-TO-C

Manual Integrations
APPROVED

MMDadoda
2/28/2020 12:33:44 PM

Quant Time: Feb 28 03:51:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

