

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022720\
 Data File : VL034704.D
 Acq On : 27 Feb 2020 10:21
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 28 01:18:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 01:15:02 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.03
2 T	Dichlorodifluoromethane	1.578	1.255	20.5	92	0.02
3	Chlorodifluoromethane	1.079	1.047	3.0	93	0.02
4	Chloromethane	0.604	0.576	4.6	89	0.02
5 T	Vinyl Chloride	0.593	0.579	2.4	92	0.02
6 T	Bromomethane	0.347	0.347	0.0	93	0.02
7	Chloroethane	0.220	0.211	4.1	88	0.02
8 T	Dichlorotetrafluoroethane	1.395	1.300	6.8	92	0.02
9 T	Propene	0.451	0.433	4.0	88	0.02
10 T	Heptane	1.161	1.195	-2.9	91	0.03
11 T	Trichlorofluoromethane	1.382	1.347	2.5	95	0.02
12 T	1,1,2-Trichlorotrifluoroeth	0.974	0.930	4.5	94	0.02
13	Ethanol	0.143	0.137	4.2	93	0.02
14 T	Bromoethene	0.406	0.402	1.0	92	0.02
15 T	Acetone	1.140	1.019	10.6	92	0.02
16 T	1,3-Butadiene	0.527	0.518	1.7	90	0.02
17	tert-Butyl alcohol	0.873	1.061	-21.5	96	0.02
18 T	1,1-Dichloroethene	0.437	0.428	2.1	93	0.02
19 T	Isopropyl Alcohol	0.679	0.802	-18.1	95	0.02
20 T	Methylene Chloride	0.424	0.377	11.1	94	0.02
21 T	Allyl Chloride	0.754	0.757	-0.4	91	0.02
22 T	trans-1,2-Dichloroethene	0.426	0.424	0.5	91	0.02
23 T	Vinyl Acetate	1.549	1.662	-7.3	92	0.02
24 T	1,1-Dichloroethane	1.237	1.212	2.0	93	0.02
25 T	Ethyl Acetate	2.017	2.013	0.2	92	0.03
26 T	Hexane	0.930	0.910	2.2	91	0.03
27 T	Carbon Disulfide	1.055	1.101	-4.4	90	0.02
28 T	Methyl tert-Butyl Ether	1.328	1.366	-2.9	91	0.03
29 T	Chloroform	1.481	1.435	3.1	93	0.02
30 T	Cyclohexane	0.721	0.737	-2.2	90	0.02
31 T	cis-1,2-Dichloroethene	0.915	0.943	-3.1	90	0.02
32 T	1,1,1-Trichloroethane	1.383	1.389	-0.4	95	0.02
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.03
34 T	2-Butanone	0.629	0.634	-0.8	91	0.02
35 T	Carbon Tetrachloride	0.670	0.684	-2.1	93	0.03
36 T	Benzene	0.864	0.862	0.2	90	0.03
37 T	1,2-Dichloroethane	0.544	0.560	-2.9	95	0.02
38 T	Trichloroethene	0.333	0.326	2.1	92	0.03
39 T	1,2-Dichloropropane	0.344	0.345	-0.3	90	0.03
40 T	1,4-Dioxane	0.117	0.126	-7.7	89	0.02
41 T	Tetrahydrofuran	0.332	0.357	-7.5	90	0.02
42 T	Bromodichloromethane	0.713	0.743	-4.2	93	0.02
43	Methyl Methacrylate	0.363	0.401	-10.5	91	0.03
44 T	2,2,4-Trimethylpentane	1.509	1.504	0.3	91	0.03
45 T	t-1,3-Dichloropropene	0.396	0.472	-19.2	90	0.03

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022720\
 Data File : VL034704.D
 Acq On : 27 Feb 2020 10:21
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 28 01:18:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 01:15:02 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.475	0.546	-14.9	91	0.03
47 T	1,1,2-Trichloroethane	0.353	0.353	0.0	91	0.02
48 T	Dibromochloromethane	0.548	0.591	-7.8	91	0.03
49 T	Bromoform	0.475	0.523	-10.1	92	0.03
50 T	4-Methyl-2-Pentanone	0.892	0.951	-6.6	92	0.03
51 T	2-Hexanone	0.702	0.784	-11.7	91	0.03
52 T	Tetrachloroethene	0.317	0.297	6.3	92	0.03
53 T	Toluene	0.919	0.981	-6.7	90	0.03
54 T	1,2-Dibromoethane	0.526	0.547	-4.0	92	0.03
55 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.03
56	1,1,1,2-Tetrachloroethane	0.426	0.404	5.2	95	0.03
57 T	Chlorobenzene	0.739	0.665	10.0	91	0.03
58 T	Ethyl Benzene	1.210	1.203	0.6	91	0.03
59 T	m/p-Xylene	1.055	0.991	6.1	90	0.03
60 T	o-Xylene	1.047	0.984	6.0	92	0.03
61 T	Styrene	0.673	0.697	-3.6	89	0.03
62	Isopropylbenzene	1.528	1.432	6.3	92	0.03
63 T	1,1,2,2-Tetrachloroethane	0.816	0.704	13.7	92	0.03
64	n-propylbenzene	0.394	0.370	6.1	89	0.03
65	tert-Butylbenzene	1.319	1.244	5.7	92	0.03
66 T	Benzyl Chloride	0.492	0.569	-15.7	91	0.03
67	sec-Butylbenzene	1.898	1.772	6.6	92	0.03
68 S	1-Bromo-4-Fluorobenzene	0.820	0.790	3.7	90	0.03
69	p-Isopropyltoluene	1.562	1.482	5.1	93	0.03
70	n-Butylbenzene	1.701	1.598	6.1	93	0.03
71	2-Chlorotoluene	1.183	1.139	3.7	92	0.03
72 T	4-Ethyltoluene	1.165	1.158	0.6	92	0.03
73 T	1,3,5-Trimethylbenzene	1.121	1.087	3.0	92	0.03
74 T	1,2,4-Trimethylbenzene	1.238	1.119	9.6	92	0.03
75 T	1,3-Dichlorobenzene	0.715	0.641	10.3	92	0.03
76 T	1,4-Dichlorobenzene	0.715	0.641	10.3	90	0.03
77 T	1,2-Dichlorobenzene	0.691	0.617	10.7	90	0.03
78 T	Hexachloro-1,3-Butadiene	0.602	0.536	11.0	96	0.03
79 T	Naphthalene	1.090	1.052	3.5	90	0.04
80 T	Naphthalene,2-methyl-	0.336	0.476	-41.7#	150	0.03
81 T	1,2,4-Trichlorobenzene	0.650	0.602	7.4	91	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0