

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022820\
 Data File : VL034750.D
 Acq On : 28 Feb 2020 23:47
 Operator : SY/AP
 Sample : L1661-08DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-8-COUNCELOR-ROOMDL

Manual Integrations
 APPROVED

MMDadoda
 3/2/2020 1:02:10 PM

Quant Time: Mar 02 12:22:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1033789	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2121611	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2083522	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1752065	10.26	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	228306	15.497	ppbv	96
15) Acetone	3.83	43	242947	2.062	ppbv	97
19) Isopropyl Alcohol	3.95	45	392282	5.591	ppbv	100
38) Trichloroethene	7.85	130	76374m	1.081	ppbv	
53) Toluene	9.83	91	42631m	0.219	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022820\
Data File : VL034750.D
Acq On : 28 Feb 2020 23:47
Operator : SY/AP
Sample : L1661-08DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-8-COUNCELOR-ROOMDL

Manual Integrations
APPROVED

MMDadoda
3/2/2020 1:02:10 PM

Quant Time: Mar 02 12:22:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

