

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022820\
 Data File : VL034754.D
 Acq On : 29 Feb 2020 2:27
 Operator : SY/AP
 Sample : L1698-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-8-4DL

Manual Integrations
 APPROVED

MMDadoda
 3/2/2020 1:02:19 PM

Quant Time: Mar 02 12:32:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1019523	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2138966	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2155859	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1777688	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	21373	0.465	ppbv	88
13) Ethanol	3.59	45	69852	4.808	ppbv	95
15) Acetone	3.84	43	159078	1.369	ppbv #	76
19) Isopropyl Alcohol	3.96	45	91885	1.328	ppbv #	94
52) Tetrachloroethene	11.25	164	132440	1.956	ppbv	90
53) Toluene	9.83	91	103445	0.526	ppbv	94
59) m/p-Xylene	13.01	91	66478m	0.292	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022820\
Data File : VL034754.D
Acq On : 29 Feb 2020 2:27
Operator : SY/AP
Sample : L1698-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-8-4DL

Manual Integrations
APPROVED

MMDadoda
3/2/2020 1:02:19 PM

Quant Time: Mar 02 12:32:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

