

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022820\
 Data File : VL034755.D
 Acq On : 29 Feb 2020 3:05
 Operator : SY/AP
 Sample : L1698-01DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-8-4DL2

Manual Integrations
 APPROVED

MMDadoda
 3/2/2020 1:02:20 PM

Quant Time: Mar 02 08:42:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	962080	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2035838	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	1999352	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1695051	10.34	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	13032m	0.951	ppbv	
15) Acetone	3.85	43	39995	0.365	ppbv	# 75
19) Isopropyl Alcohol	3.98	45	17298m	0.265	ppbv	
20) Methylene Chloride	4.33	84	8433	0.207	ppbv	85
52) Tetrachloroethene	11.25	164	20483m	0.318	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022820\
Data File : VL034755.D
Acq On : 29 Feb 2020 3:05
Operator : SY/AP
Sample : L1698-01DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-8-4DL2

Manual Integrations
APPROVED

MMDadoda
3/2/2020 1:02:20 PM

Quant Time: Mar 02 08:42:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

