

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL030121\
 Data File : VL036414.D
 Acq On : 1 Mar 2021 14:39
 Operator : SY/AP
 Sample : M1463-14DL DUP 4X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NS-6DL DUP

Manual Integrations
 APPROVED

MMDadoda
 3/2/2021 3:30:08 PM

Quant Time: Mar 02 11:39:55 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 16:04:02 2021
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1569102	10.00	ppbv	0.04
33) 1,4-Difluorobenzene	7.20	114	3314341	10.00	ppbv	0.05
55) Chlorobenzene-d5	12.12	117	2915099	10.00	ppbv	0.06

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2234706	9.32	ppbv	0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Ethanol	3.64	45	69676m	5.103	ppbv	
20) Methylene Chloride	4.37	84	499203	8.067	ppbv	94
26) Hexane	5.71	57	595299	3.913	ppbv #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL030121\
Data File : VL036414.D
Acq On : 1 Mar 2021 14:39
Operator : SY/AP
Sample : M1463-14DLUP 4X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NS-6DLUP

Manual Integrations
APPROVED

MMDadoda
3/2/2021 3:30:08 PM

Quant Time: Mar 02 11:39:55 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

