

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL030121\
Data File : VL036421.D
Acq On : 1 Mar 2021 19:08
Operator : SY/AP
Sample : QC022321 CAN 10727
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC022321 CAN 10727

Quant Time: Mar 02 08:55:50 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10 16:04:02 2021
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1666177	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.19	114	3582365	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.10	117	3258617	10.00	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2610027	9.73	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.30%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL030121\
Data File : VL036421.D
Acq On : 1 Mar 2021 19:08
Operator : SY/AP
Sample : QC022321 CAN 10727
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC022321 CAN 10727

Quant Time: Mar 02 08:55:50 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 16:04:02 2021
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

