

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
 Data File : VL038454.D
 Acq On : 1 Mar 2022 17:09
 Operator : SY/AP
 Sample : N1720-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/02/2022
 Supervised By :Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:06:49 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 2022

QLast Update : Tue Feb 22 19:45:51 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	951589	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.176	114	2361593	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.082	117	2064333	10.000	ppbv	# 0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.413	95	1617498	9.470	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.700%
Target Compounds						
						Qvalue
9) Propene	3.031	41	17980m	0.324	ppbv	
10) Heptane	8.127	43	25407m	0.188	ppbv	
13) Ethanol	3.636	45	8591m	1.552	ppbv	
15) Acetone	3.886	43	67074m	0.829	ppbv	
20) Methylene Chloride	4.353	84	17356	0.433	ppbv	87
26) Hexane	5.693	57	271060	2.753	ppbv #	29
29) Chloroform	5.755	83	66520	0.414	ppbv	99
34) 2-Butanone	5.279	43	17702	0.238	ppbv #	76
52) Tetrachloroethene	11.217	164	366434m	5.309	ppbv	
62) Isopropylbenzene	14.664	105	33253	0.108	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
Data File : VL038454.D
Acq On : 1 Mar 2022 17:09
Operator : SY/AP
Sample : N1720-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/02/2022
Supervised By : Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:06:49 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

