

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
 Data File : VL038457.D
 Acq On : 1 Mar 2022 19:09
 Operator : SY/AP
 Sample : N1720-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/02/2022
 Supervised By :Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:07:57 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 2022

QLast Update : Tue Feb 22 19:45:51 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	892006	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.176	114	2152917	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.085	117	1825187	10.000	ppbv	# 0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.410	95	1370477	9.075	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	90.700%
Target Compounds						
					Qvalue	
9) Propene	3.031	41	9546m	0.184	ppbv	
13) Ethanol	3.629	45	106702m	20.558	ppbv	
15) Acetone	3.874	43	31506m	0.415	ppbv	
19) Isopropyl Alcohol	3.996	45	26729m	0.608	ppbv	
20) Methylene Chloride	4.349	84	15885	0.423	ppbv	90
26) Hexane	5.690	57	12519m	0.136	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
Data File : VL038457.D
Acq On : 1 Mar 2022 19:09
Operator : SY/AP
Sample : N1720-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/02/2022
Supervised By : Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:07:57 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

