

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
 Data File : VL038459.D
 Acq On : 1 Mar 2022 20:29
 Operator : SY/AP
 Sample : N1720-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/02/2022
 Supervised By :Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:08:31 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 2022
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.674	49	863819	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.179	114	2016724	10.000	ppbv	0.03
55) Chlorobenzene-d5	12.085	117	1684992	10.000	ppbv	# 0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.417	95	1272626	9.128	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.300%
Target Compounds						
					Qvalue	
9) Propene	3.031	41	9774m	0.194	ppbv	
13) Ethanol	3.636	45	24619m	4.898	ppbv	
15) Acetone	3.880	43	30892m	0.421	ppbv	
19) Isopropyl Alcohol	4.002	45	15951m	0.375	ppbv	
20) Methylene Chloride	4.356	84	21344m	0.586	ppbv	
26) Hexane	5.693	57	11995	0.134	ppbv	# 45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030122\
Data File : VL038459.D
Acq On : 1 Mar 2022 20:29
Operator : SY/AP
Sample : N1720-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/02/2022
Supervised By : Mahesh Dadoda 03/02/2022

Quant Time: Mar 02 03:08:31 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

