


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL030122\  
Data File : VL038443.D  
Acq On : 1 Mar 2022 8:40  
Operator : SY/AP  
Sample : BFB  
Misc : 400mL/MSVOA\_L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_L\METHODS\VL022222AIR.M  
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Tue Feb 22 19:45:51 2022  
Last Update : Tue Feb 22 19:45:51 2022



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3582

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw    | Result    |
|--------|---------|--------|--------|-------|--------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn    | Pass/Fail |
| 50     | 95      | 8      | 40     | 24.5  | 125691 | PASS      |
| 75     | 95      | 30     | 66     | 57.0  | 292651 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 513152 | PASS      |
| 96     | 95      | 5      | 9      | 6.8   | 34688  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.1   | 3230   | PASS      |
| 174    | 95      | 50     | 120    | 58.4  | 299925 | PASS      |
| 175    | 174     | 4      | 9      | 7.4   | 22091  | PASS      |
| 176    | 174     | 93     | 101    | 97.0  | 291051 | PASS      |
| 177    | 176     | 5      | 9      | 6.4   | 18488  | PASS      |

Average of 14.388 to 14.394 min.: VL038443.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 36.00 | 4817   | 50.00 | 125691 | 60.95 | 25461  | 72.00 | 2604   |
| 37.00 | 29699  | 51.00 | 39867  | 62.00 | 25067  | 73.00 | 2197   |
| 38.00 | 27525  | 51.95 | 2074   | 63.05 | 19904  | 74.00 | 9556   |
| 39.00 | 10900  | 52.90 | 66     | 63.95 | 1735   | 75.00 | 29265  |
| 42.95 | 227    | 53.80 | 150    | 64.95 | 842    | 76.00 | 23909  |
| 44.00 | 2795   | 54.90 | 1010   | 66.95 | 1473   | 76.95 | 2639   |
| 45.00 | 5617   | 55.10 | 575    | 68.00 | 63536  | 77.90 | 1069   |
| 46.15 | 205    | 56.00 | 8090   | 69.00 | 62661  | 78.10 | 509    |
| 47.00 | 5408   | 57.00 | 17280  | 70.00 | 5138   | 78.90 | 13625  |
| 47.90 | 3297   | 58.10 | 487    | 70.85 | 271    | 79.95 | 3557   |
| 49.00 | 22411  | 59.95 | 4722   | 71.20 | 177    | 80.90 | 14191  |

Average of 14.388 to 14.394 min.: VL038443.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|--------|--------|--------|--------|--------|--------|
| 82.00 | 2473   | 92.95  | 23859  | 106.90 | 328    | 115.90 | 1955   |
| 82.85 | 506    | 94.00  | 63443  | 108.20 | 76     | 116.85 | 2866   |
| 83.80 | 50     | 95.00  | 513152 | 108.60 | 85     | 117.95 | 1819   |
| 85.70 | 111    | 96.00  | 34688  | 109.80 | 123    | 118.90 | 2445   |
| 86.05 | 276    | 96.85  | 619    | 110.50 | 109    | 122.90 | 279    |
| 86.95 | 14258  | 97.10  | 202    | 111.00 | 516    | 123.50 | 137    |
| 87.95 | 14848  | 102.85 | 105    | 111.95 | 259    | 123.80 | 115    |
| 89.30 | 54     | 103.85 | 2126   | 112.70 | 82     | 124.00 | 74     |
| 90.80 | 812    | 104.85 | 749    | 113.05 | 285    | 125.95 | 243    |
| 90.95 | 1224   | 105.85 | 2010   | 114.60 | 57     | 126.90 | 129    |
| 92.00 | 13714  | 106.70 | 120    | 114.95 | 377    | 127.85 | 1435   |

Average of 14.388 to 14.394 min.: VL038443.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 128.80 | 822    | 140.95 | 4299   | 148.95 | 426    | 157.80 | 118    |
| 129.95 | 1420   | 141.80 | 423    | 149.80 | 86     | 158.90 | 571    |
| 130.85 | 599    | 142.10 | 168    | 150.00 | 318    | 160.95 | 465    |
| 133.80 | 78     | 142.85 | 4913   | 151.80 | 103    | 161.70 | 102    |
| 134.90 | 240    | 143.90 | 148    | 152.60 | 108    | 166.80 | 51     |
| 135.05 | 653    | 144.85 | 462    | 152.85 | 199    | 168.70 | 144    |
| 135.80 | 54     | 145.65 | 293    | 153.75 | 253    | 169.10 | 53     |
| 136.95 | 775    | 145.90 | 193    | 154.80 | 357    | 169.30 | 99     |
| 138.90 | 90     | 146.85 | 230    | 155.00 | 714    | 169.85 | 169    |
| 139.80 | 299    | 147.90 | 1109   | 155.85 | 166    | 170.30 | 156    |
| 140.00 | 147    | 148.70 | 64     | 156.95 | 707    | 170.60 | 266    |

Average of 14.388 to 14.394 min.: VL038443.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z | abund. | m/z | abund. |
|--------|--------|--------|--------|-----|--------|-----|--------|
| 171.10 | 401    | 227.10 | 60     |     |        |     |        |
| 171.85 | 844    |        |        |     |        |     |        |
| 172.10 | 572    |        |        |     |        |     |        |
| 172.95 | 3230   |        |        |     |        |     |        |
| 173.90 | 299925 |        |        |     |        |     |        |
| 174.95 | 22091  |        |        |     |        |     |        |
| 175.90 | 291051 |        |        |     |        |     |        |
| 176.95 | 18488  |        |        |     |        |     |        |
| 178.00 | 437    |        |        |     |        |     |        |
| 193.40 | 111    |        |        |     |        |     |        |
| 197.60 | 72     |        |        |     |        |     |        |

Instrument :

MSVOA\_L

ClientSampleId :

BFB