

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030222\  
Data File : VL038476.D  
Acq On : 2 Mar 2022 11:09  
Operator : SY/AP  
Sample : VL0302ABL01  
Misc : 400mL/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
VL0302ABL01

Quant Time: Mar 02 22:50:12 2022  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M  
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA\_L Tue Feb 22 19:45:51 2022  
QLast Update : Tue Feb 22 19:45:51 2022  
Response via : Initial Calibration

| Internal Standards      | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------|-------|------|----------|-------|-------|----------|
| 1) Bromochloromethane   | 5.66  | 49   | 914756   | 10.00 | ppbv  | 0.02     |
| 33) 1,4-Difluorobenzene | 7.17  | 114  | 2208102  | 10.00 | ppbv  | 0.02     |
| 55) Chlorobenzene-d5    | 12.08 | 117  | 1871637  | 10.00 | ppbv  | 0.02     |

System Monitoring Compounds

|                             |        |       |          |          |      |        |
|-----------------------------|--------|-------|----------|----------|------|--------|
| 68) 1-Bromo-4-Fluorobenzene | 14.41  | 95    | 1523916  | 9.84     | ppbv | 0.02   |
| Spiked Amount               | 10.000 | Range | 65 - 135 | Recovery | =    | 98.40% |

| Target Compounds | Ovalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030222\  
Data File : VL038476.D  
Acq On : 2 Mar 2022 11:09  
Operator : SY/AP  
Sample : VL0302ABL01  
Misc : 400mL/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
VL0302ABL01

Quant Time: Mar 02 22:50:12 2022  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Tue Feb 22 19:45:51 2022  
QLast Update : Tue Feb 22 19:45:51 2022  
Response via : Initial Calibration

