

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030222\
Data File : VL038478.D
Acq On : 2 Mar 2022 12:35
Operator : SY/AP
Sample : QC030122 CAN 10118
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC030122 CAN 10118

Quant Time: Mar 03 11:53:25 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	950212	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.17	114	2234737	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.08	117	1863654	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1552476	10.07	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030222\
Data File : VL038478.D
Acq On : 2 Mar 2022 12:35
Operator : SY/AP
Sample : OC030122 CAN 10118
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC030122 CAN 10118

Quant Time: Mar 03 11:53:25 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

