

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030222\
 Data File : VL038481.D
 Acq On : 2 Mar 2022 14:28
 Operator : SY/AP
 Sample : N1723-03DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 FSG1DL2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/03/2022
 Supervised By : Mahesh Dadoda 03/07/2022

Quant Time: Mar 05 16:16:21 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	904764	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.17	114	2126984	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.08	117	1925448	10.00	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1580610	9.92	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	184975	3.507	ppbv #	69
15) Acetone	3.85	43	779857	10.136	ppbv	99
30) Cyclohexane	7.14	84	17631m	0.230	ppbv	
36) Benzene	6.89	78	61450	0.328	ppbv #	93
44) 2,2,4-Trimethylpentane	7.87	57	68420m	0.213	ppbv	
53) Toluene	9.81	91	124150	0.611	ppbv	98
62) Isopropylbenzene	14.67	105	64085	0.223	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030222\
Data File : VL038481.D
Acq On : 2 Mar 2022 14:28
Operator : SY/AP
Sample : N1723-03DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
FSG1DL2

Manual Integrations
APPROVED

Quant Time: Mar 05 16:16:21 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222\
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/03/2022
Supervised By :Mahesh Dadoda 03/07/2022

