

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030222\
 Data File : VL038495.D
 Acq On : 2 Mar 2022 23:29
 Operator : SY/AP
 Sample : CAN 10454
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10454

Quant Time: Mar 03 10:23:21 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	830963	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.18	114	1963975	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.09	117	1704620	10.00	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1373193	9.74	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030222\
Data File : VL038495.D
Acq On : 2 Mar 2022 23:29
Operator : SY/AP
Sample : CAN 10454
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10454

Quant Time: Mar 03 10:23:21 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

