

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030222\
 Data File : VL038495.D
 Acq On : 2 Mar 2022 23:29
 Operator : SY/AP
 Sample : CAN 10454
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10454

Quant Time: Mar 03 10:23:21 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	830963	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.18	114	1963975	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.09	117	1704620	10.00	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1373193	9.74	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.40%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030222\
Data File : VL038495.D
Acq On : 2 Mar 2022 23:29
Operator : SY/AP
Sample : CAN 10454
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10454

Quant Time: Mar 03 10:23:21 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.67 Instrument:

Delta R.T. MSVOA_L

Lab File: ClientSampleId:

Acq: 2 Mar 2022 23:29


Tot Ion: 49 Resp: 830963

Ion Ratio Lower Upper

49 100

128 45.2 24.3 72.9




130 60.5 30.1 90.5



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): V

Ion 130.00 (129.70 to 130.70): V



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.18 min Scan# 1367

Delta R.T. 0.03 min

Lab File: VL038495.D

Acq: 2 Mar 2022 23:29


Tgt Ion:114 Resp: 1963975

Ion Ratio Lower Upper

114 100

63 28.0 22.2 33.2



88 18.8 15.2 22.8




Abundance Ion 114.10 (113.80 to 114.80): V

Ion 63.10 (62.80 to 63.80): VL0





Ion 88.10 (87.80 to 88.80): VL0





#55
 Chlorobenzene-d5
 Concen: 10.000 ppbv
 RT: 12.09 min
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampleId:
 Acq: 2 Mar 2022 23:29

Tot Ion: 117 Resp: 1704620
 Ion Ratio Lower Upper
 117 100
 82 74.1 54.9 82.3
 119 34.0 46.1 69.1#



#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.736 ppbv
 RT: 14.42 min Scan# 3619
 Delta R.T.: 0.04 min
 Lab File: VL038495.D
 Acq: 2 Mar 2022 23:29

Tgt Ion: 95 Resp: 1373193
 Ion Ratio Lower Upper
 95 100
 174 62.1 49.0 73.4
 175 4.4 3.6 5.4

