

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030320\
 Data File : VL034765.D
 Acq On : 3 Mar 2020 15:48
 Operator : SY/AP
 Sample : L1704-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-3-4DL

Manual Integrations
 APPROVED

MMDadoda
 3/4/2020 4:49:08 PM

Quant Time: Mar 04 11:14:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1047913	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.20	114	2106096	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.14	117	2091686	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.48	95	1784290	10.40	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	107204m	2.268	ppbv	
13) Ethanol	3.61	45	39233	2.627	ppbv	99
15) Acetone	3.84	43	2164590	18.127	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030320\
Data File : VL034765.D
Acq On : 3 Mar 2020 15:48
Operator : SY/AP
Sample : L1704-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-3-4DL

Manual Integrations
APPROVED

MMDadoda
3/4/2020 4:49:08 PM

Quant Time: Mar 04 11:14:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

