

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030320\
 Data File : VL034773.D
 Acq On : 3 Mar 2020 21:08
 Operator : SY/AP
 Sample : L1698-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-7-4DL

Manual Integrations
 APPROVED

MMDadoda
 3/4/2020 4:49:28 PM

Quant Time: Mar 04 08:01:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb 28 2020
 QLast Update : Fri Feb 28 01:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1174066	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.21	114	2417351	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.15	117	2427727	10.00	ppbv	0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.49	95	2070571	10.40	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
11) Trichlorofluoromethane	3.94	101	130408	0.804	ppbv	97
13) Ethanol	3.61	45	4586	0.274	ppbv #	1
15) Acetone	3.85	43	305511	2.284	ppbv	95
20) Methylene Chloride	4.35	84	31492	0.633	ppbv	82
52) Tetrachloroethene	11.28	164	62935	0.822	ppbv	87
53) Toluene	9.86	91	105064m	0.473	ppbv	
59) m/p-Xylene	13.02	91	47010m	0.184	ppbv	
74) 1,2,4-Trimethylbenzene	16.84	105	37126m	0.124	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030320\
Data File : VL034773.D
Acq On : 3 Mar 2020 21:08
Operator : SY/AP
Sample : L1698-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-7-4DL

Manual Integrations
APPROVED

MMDadoda
3/4/2020 4:49:28 PM

Quant Time: Mar 04 08:01:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Feb
QLast Update : Fri Feb 28 01:15:02 2020
Response via : Initial Calibration

