

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030420\
 Data File : VL034781.D
 Acq On : 4 Mar 2020 4:34
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

apatel
 3/4/2020 4:45:23 PM

Quant Time: Mar 04 06:33:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 04 2020
 QLast Update : Wed Mar 04 05:54:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1025984	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2098844	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.13	117	2049504	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1757491	10.46	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.23	62	2672	0.044	ppbv	# 80
32) 1,1,1-Trichloroethane	6.52	97	6238m	0.044	ppbv	
35) Carbon Tetrachloride	7.04	117	6368m	0.045	ppbv	
38) Trichloroethene	7.86	130	2741m	0.039	ppbv	
52) Tetrachloroethene	11.26	164	2385m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.73	83	6088m	0.036	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030420\
Data File : VL034781.D
Acq On : 4 Mar 2020 4:34
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

apatel
3/4/2020 4:45:23 PM

Quant Time: Mar 04 06:33:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 04 05:54:58 2020
Response via : Initial Calibration

