

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030420\
 Data File : VL034784.D
 Acq On : 4 Mar 2020 7:13
 Operator : SY/AP
 Sample : VL0304ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0304ABL01

Manual Integrations
 APPROVED

apatel
 3/4/2020 4:45:30 PM

Quant Time: Mar 04 07:58:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Mar 04 16:21:26 2020
 QLast Update : Wed Mar 04 06:36:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	997791	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.19	114	2018252	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.12	117	2052330	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	1709548	9.99	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	26108m	0.219	ppbv	
20) Methylene Chloride	4.33	84	10660m	0.250	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030420\
Data File : VL034784.D
Acq On : 4 Mar 2020 7:13
Operator : SY/AP
Sample : VL0304ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0304ABL01

Manual Integrations
APPROVED

apatel
3/4/2020 4:45:30 PM

Quant Time: Mar 04 07:58:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 04 06:36:39 2020
Response via : Initial Calibration

