


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL030420\
Data File : VL034775.D
Acq On : 4 Mar 2020 00:45
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030420AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Mar 04 06:36:39 2020
Last Update : Wed Mar 04 06:36:39 2020



AutoFind: Scans 3635, 3636, 3637; Background Corrected with Scan 3612

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.7	206186	PASS
75	95	30	66	58.9	492458	PASS
95	95	100	100	100.0	835413	PASS
96	95	5	9	7.1	58965	PASS
173	174	0.00	2	0.3	1601	PASS
174	95	50	120	64.6	539456	PASS
175	174	4	9	7.4	39960	PASS
176	174	93	101	95.7	516032	PASS
177	176	5	9	6.3	32320	PASS