

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030525\
 Data File : VL042108.D
 Acq On : 05 Mar 2025 17:31
 Operator : SY/MD
 Sample : Q1483-06 1200X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/06/2025
 Supervised By :Mahesh Dadoda 03/06/2025

Quant Time: Mar 06 01:10:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Feb 26 01:08:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.810	49	157233	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.991	114	350842	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.914	117	294774	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.403	95	207279	9.140	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.400%
Target Compounds						
19) Isopropyl Alcohol	1.900	45	4948	0.620	ppbv	# 40
22) trans-1,2-Dichloroethene	2.356	96	819m	0.119	ppbv	
31) cis-1,2-Dichloroethene	2.742	61	20823	0.979	ppbv	99
38) Trichloroethene	4.583	130	30337m	2.023	ppbv	
52) Tetrachloroethene	8.267	164	1666581	132.549	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030525\
Data File : VL042108.D
Acq On : 05 Mar 2025 17:31
Operator : SY/MD
Sample : Q1483-06 1200X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV4

Quant Time: Mar 06 01:10:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Feb 26 01:08:36 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/06/2025
Supervised By : Mahesh Dadoda 03/06/2025

