

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030620\
 Data File : VL034820.D
 Acq On : 6 Mar 2020 17:55
 Operator : SY/AP
 Sample : VL0306ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0306ABL01

Manual Integrations
 APPROVED

MMDadoda
 3/9/2020 9:20:27 AM

Quant Time: Mar 07 04:24:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar 0
 QLast Update : Wed Mar 04 06:36:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	873375	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.18	114	1991922	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.12	117	1945585	10.00	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	1530790	9.44	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.40%
Target Compounds						
15) Acetone	3.85	43	18629m	0.178	ppbv	Ovalue
20) Methylene Chloride	4.32	84	5227	0.140	ppbv	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030620\
Data File : VL034820.D
Acq On : 6 Mar 2020 17:55
Operator : SY/AP
Sample : VL0306ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0306ABL01

Manual Integrations
APPROVED

MMDadoda
3/9/2020 9:20:27 AM

Quant Time: Mar 07 04:24:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar
QLast Update : Wed Mar 04 06:36:39 2020
Response via : Initial Calibration

