

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030620\
 Data File : VL034825.D
 Acq On : 6 Mar 2020 21:51
 Operator : SY/AP
 Sample : L1698-04DL 300X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-4-4DL

Manual Integrations
 APPROVED

MMDadoda
 3/9/2020 9:20:39 AM

Quant Time: Mar 07 09:11:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 07 09:11:14 2020
 QLast Update : Wed Mar 04 06:36:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	821892	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.19	114	1971169	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.12	117	1910802	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1553708	9.75	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.22	62	327160	6.140	ppbv	98
9) Propene	2.98	41	82751	2.245	ppbv	89
10) Heptane	8.16	43	2965144	31.656	ppbv	99
19) Isopropyl Alcohol	3.97	45	26610	0.451	ppbv #	93
26) Hexane	5.70	57	8343549	112.295	ppbv #	42
30) Cyclohexane	7.16	84	2320363	40.821	ppbv	97
31) cis-1,2-Dichloroethene	5.54	61	17774m	0.233	ppbv	
36) Benzene	6.91	78	3306587	19.631	ppbv	98
53) Toluene	9.84	91	993110	5.571	ppbv	99
58) Ethyl Benzene	12.75	91	68930	0.293	ppbv	96
59) m/p-Xylene	13.00	91	61280m	0.301	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030620\
Data File : VL034825.D
Acq On : 6 Mar 2020 21:51
Operator : SY/AP
Sample : L1698-04DL 300X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-4-4DL

Manual Integrations
APPROVED

MMDadoda
3/9/2020 9:20:39 AM

Quant Time: Mar 07 09:11:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 04 06:36:39 2020
Response via : Initial Calibration

