

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031020\
 Data File : VL034863.D
 Acq On : 10 Mar 2020 11:30
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 11 06:34:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	1.369	1.619	-18.3	145	0.00
3	Chlorodifluoromethane	1.117	1.048	6.2	97	0.00
4	Chloromethane	0.631	0.617	2.2	99	0.00
5 T	Vinyl Chloride	0.607	0.573	5.6	95	0.00
6 T	Bromomethane	0.353	0.345	2.3	99	0.00
7	Chloroethane	0.219	0.209	4.6	98	0.00
8 T	Dichlorotetrafluoroethane	1.422	1.299	8.6	96	0.00
9 T	Propene	0.533	0.485	9.0	95	0.00
10 T	Heptane	1.313	1.268	3.4	96	0.00
11 T	Trichlorofluoromethane	1.457	1.327	8.9	97	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.020	0.939	7.9	97	0.00
13	Ethanol	0.162	0.114	29.6	88	0.00
14 T	Bromoethene	0.440	0.415	5.7	95	0.00
15 T	Acetone	1.296	1.193	7.9	109	0.00
16 T	1,3-Butadiene	0.586	0.562	4.1	98	0.00
17	tert-Butyl alcohol	1.249	0.901	27.9	78	0.00
18 T	1,1-Dichloroethene	0.452	0.433	4.2	98	0.00
19 T	Isopropyl Alcohol	0.818	0.607	25.8	80	0.00
20 T	Methylene Chloride	0.410	0.367	10.5	95	0.00
21 T	Allyl Chloride	0.832	0.782	6.0	96	0.00
22 T	trans-1,2-Dichloroethene	0.458	0.432	5.7	97	0.00
23 T	Vinyl Acetate	1.915	1.855	3.1	98	0.00
24 T	1,1-Dichloroethane	1.283	1.209	5.8	97	0.00
25 T	Ethyl Acetate	2.309	2.130	7.8	96	0.00
26 T	Hexane	0.994	0.922	7.2	94	0.00
27 T	Carbon Disulfide	1.128	1.091	3.3	92	0.00
28 T	Methyl tert-Butyl Ether	1.654	1.498	9.4	91	0.00
29 T	Chloroform	1.517	1.406	7.3	97	0.00
30 T	Cyclohexane	0.830	0.791	4.7	95	0.00
31 T	cis-1,2-Dichloroethene	1.017	0.972	4.4	96	0.00
32 T	1,1,1-Trichloroethane	1.478	1.404	5.0	95	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
34 T	2-Butanone	0.665	0.629	5.4	97	0.00
35 T	Carbon Tetrachloride	0.622	0.638	-2.6	95	0.00
36 T	Benzene	0.891	0.838	5.9	95	0.00
37 T	1,2-Dichloroethane	0.503	0.486	3.4	96	0.00
38 T	Trichloroethene	0.339	0.319	5.9	94	0.00
39 T	1,2-Dichloropropane	0.353	0.332	5.9	95	0.00
40 T	1,4-Dioxane	0.133	0.112	15.8	90	0.00
41 T	Tetrahydrofuran	0.360	0.353	1.9	94	0.00
42 T	Bromodichloromethane	0.685	0.665	2.9	93	0.00
43	Methyl Methacrylate	0.376	0.377	-0.3	96	0.00
44 T	2,2,4-Trimethylpentane	1.506	1.443	4.2	96	0.00
45 T	t-1,3-Dichloropropene	0.441	0.451	-2.3	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031020\
 Data File : VL034863.D
 Acq On : 10 Mar 2020 11:30
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 11 06:34:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.524	0.522	0.4	93	0.00
47 T	1,1,2-Trichloroethane	0.362	0.338	6.6	92	0.00
48 T	Dibromochloromethane	0.587	0.556	5.3	88	0.00
49 T	Bromoform	0.523	0.484	7.5	83	0.00
50 T	4-Methyl-2-Pentanone	0.957	0.926	3.2	95	0.00
51 T	2-Hexanone	0.777	0.761	2.1	94	0.00
52 T	Tetrachloroethene	0.353	0.300	15.0	89	0.00
53 T	Toluene	1.029	0.980	4.8	94	0.00
54 T	1,2-Dibromoethane	0.546	0.516	5.5	92	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
56	1,1,1,2-Tetrachloroethane	0.449	0.420	6.5	88	0.00
57 T	Chlorobenzene	0.782	0.717	8.3	92	0.00
58 T	Ethyl Benzene	1.354	1.290	4.7	92	0.00
59 T	m/p-Xylene	1.135	1.047	7.8	92	0.00
60 T	o-Xylene	1.108	1.023	7.7	92	0.00
61 T	Styrene	0.777	0.767	1.3	90	0.00
62	Isopropylbenzene	1.675	1.526	8.9	90	0.00
63 T	1,1,2,2-Tetrachloroethane	0.752	0.695	7.6	90	0.00
64	n-propylbenzene	0.436	0.397	8.9	87	0.00
65	tert-Butylbenzene	1.464	1.321	9.8	89	0.00
66 T	Benzyl Chloride	0.611	0.649	-6.2	91	0.00
67	sec-Butylbenzene	2.030	1.876	7.6	90	0.00
68 S	1-Bromo-4-Fluorobenzene	0.780	0.806	-3.3	93	0.00
69	p-Isopropyltoluene	1.702	1.590	6.6	89	0.00
70	n-Butylbenzene	1.737	1.649	5.1	91	0.00
71	2-Chlorotoluene	1.268	1.156	8.8	89	0.00
72 T	4-Ethyltoluene	1.308	1.226	6.3	90	0.00
73 T	1,3,5-Trimethylbenzene	1.225	1.155	5.7	90	0.00
74 T	1,2,4-Trimethylbenzene	1.262	1.190	5.7	91	0.00
75 T	1,3-Dichlorobenzene	0.761	0.690	9.3	90	0.00
76 T	1,4-Dichlorobenzene	0.759	0.694	8.6	89	0.00
77 T	1,2-Dichlorobenzene	0.737	0.677	8.1	89	0.00
78 T	Hexachloro-1,3-Butadiene	0.616	0.632	-2.6	99	0.00
79 T	Naphthalene	0.984	1.212	-23.2	91	0.00
80 T	Naphthalene,2-methyl-	0.257	0.451	-75.5#	76	0.00
81 T	1,2,4-Trichlorobenzene	0.625	0.693	-10.9	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0