

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031224\
 Data File : VL041100.D
 Acq On : 12 Mar 2024 15:46
 Operator : SY/MD
 Sample : QC022624 CAN 10054
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC022624 CAN 10054

Quant Time: Mar 13 01:30:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Mar 13 01:22:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	1245991	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	3116396	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2698345	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.789	95	2111886	10.010	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031224\
Data File : VL041100.D
Acq On : 12 Mar 2024 15:46
Operator : SY/MD
Sample : QC022624 CAN 10054
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC022624 CAN 10054

Quant Time: Mar 13 01:30:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Mar 13 01:22:02 2024
Response via : Initial Calibration









#1
 Bromochloromethane
 Concen: 10.000 ppbv
 RT: 5.201 min Scan# 808
 Delta R.T. 0.009 min
 Lab File: VL041100.D
 Acq: 12 Mar 2024 15:46




Instrument :
 MSVOA_L
 ClientSampleId :
 QC022624 CAN 10054


Tgt Ion: 49 Resp: 1245991
 Ion Ratio Lower Upper
 49 100
 128 46.9 23.8 71.5
 130 58.8 30.3 90.9



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 6.661 min Scan# 1262
 Delta R.T. 0.006 min
 Lab File: VL041100.D
 Acq: 12 Mar 2024 15:46

Tgt Ion: 114 Resp: 3116396
 Ion Ratio Lower Upper
 114 100
 63 26.3 21.0 31.6
 88 18.0 14.6 22.0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 11.481 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL041100.D
Acq: 12 Mar 2024 15:46

Instrument :
MSVOA_L
ClientSampleId :
QC022624 CAN 10054

Tgt Ion:117 Resp: 2698345
Ion Ratio Lower Upper
117 100
82 68.4 51.4 77.2
119 33.5 23.7 35.5



#68
1-Bromo-4-Fluorobenzene
Concen: 10.010 ppbv
RT: 13.789 min Scan# 3479
Delta R.T. 0.006 min
Lab File: VL041100.D
Acq: 12 Mar 2024 15:46

Tgt Ion: 95 Resp: 2111886
Ion Ratio Lower Upper
95 100
174 62.0 47.8 71.8
175 4.6 3.5 5.3

