

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031319\
 Data File : VL033304.D
 Acq On : 13 Mar 2019 19:43
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 3/15/2019 10:41:50 PM

Quant Time: Mar 13 20:28:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031319AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 14 00:00:00 2019
 QLast Update : Wed Mar 13 17:31:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	759169	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1685236	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1889779	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1476872	10.12	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.29	62	2053m	0.039	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	4797m	0.038	ppbv	
35) Carbon Tetrachloride	7.13	117	4672m	0.038	ppbv	
38) Trichloroethene	7.96	130	2296m	0.038	ppbv	
52) Tetrachloroethene	11.37	164	2598m	0.046	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.83	83	6078m	0.043	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL031319\
Data File : VL033304.D
Acq On : 13 Mar 2019 19:43
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
3/15/2019 10:41:50 PM

Quant Time: Mar 13 20:28:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL031319AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 13 17:31:13 2019
Response via : Initial Calibration

