

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031319\
 Data File : VL033307.D
 Acq On : 14 Mar 2019 2:34
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL031319

Quant Time: Mar 15 02:11:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031319AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 14 00:01:31 2019
 QLast Update : Thu Mar 14 00:01:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	1.850	1.510	18.4	76	0.00
3	Chlorodifluoromethane	1.304	1.192	8.6	98	0.00
4	Chloromethane	0.706	0.637	9.8	97	0.00
5 T	Vinyl Chloride	0.707	0.630	10.9	99	0.00
6 T	Bromomethane	0.426	0.403	5.4	99	0.00
7	Chloroethane	0.253	0.248	2.0	101	0.00
8 T	Dichlorotetrafluoroethane	1.690	1.548	8.4	100	0.00
9 T	Propene	0.589	0.544	7.6	98	0.00
10 T	Heptane	1.439	1.347	6.4	97	0.00
11 T	Trichlorofluoromethane	1.859	1.722	7.4	99	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.253	1.145	8.6	100	0.00
13	Ethanol	0.180	0.128	28.9	107	0.00
14 T	Bromoethene	0.521	0.512	1.7	103	0.00
15 T	Acetone	1.375	1.110	19.3	102	0.00
16 T	1,3-Butadiene	0.563	0.509	9.6	92	0.00
17	tert-Butyl alcohol	0.240	0.213	11.2	95	0.00
18 T	1,1-Dichloroethene	0.554	0.495	10.6	92	0.00
19 T	Isopropyl Alcohol	0.841	0.764	9.2	103	0.00
20 T	Methylene Chloride	0.557	0.435	21.9	97	0.00
21 T	Allyl Chloride	0.830	0.801	3.5	98	0.00
22 T	trans-1,2-Dichloroethene	0.558	0.517	7.3	95	0.00
23 T	Vinyl Acetate	1.996	1.924	3.6	95	0.00
24 T	1,1-Dichloroethane	1.444	1.308	9.4	92	0.00
25 T	Ethyl Acetate	2.505	2.278	9.1	97	0.00
26 T	Hexane	1.170	1.013	13.4	98	0.00
27 T	Carbon Disulfide	1.252	1.296	-3.5	98	0.00
28 T	Methyl tert-Butyl Ether	1.682	1.597	5.1	92	0.00
29 T	Chloroform	1.819	1.679	7.7	99	0.00
30 T	Cyclohexane	0.909	0.859	5.5	96	0.00
31 T	cis-1,2-Dichloroethene	1.127	1.053	6.6	93	0.00
32 T	1,1,1-Trichloroethane	1.766	1.656	6.2	100	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
34 T	2-Butanone	0.691	0.620	10.3	92	0.00
35 T	Carbon Tetrachloride	0.769	0.724	5.9	98	0.00
36 T	Benzene	0.982	0.902	8.1	96	0.00
37 T	1,2-Dichloroethane	0.587	0.541	7.8	100	0.00
38 T	Trichloroethene	0.388	0.337	13.1	97	0.00
39 T	1,2-Dichloropropane	0.377	0.344	8.8	97	0.00
40 T	1,4-Dioxane	0.136	0.123	9.6	96	0.00
41 T	Tetrahydrofuran	0.384	0.361	6.0	97	0.00
42 T	Bromodichloromethane	0.822	0.780	5.1	97	0.00
43	Methyl Methacrylate	0.391	0.366	6.4	95	0.00
44 T	2,2,4-Trimethylpentane	1.679	1.496	10.9	97	0.00
45 T	t-1,3-Dichloropropene	0.459	0.490	-6.8	95	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031319\
 Data File : VL033307.D
 Acq On : 14 Mar 2019 2:34
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL031319

Quant Time: Mar 15 02:11:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031319AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 14 00:01:31 2019
 QLast Update : Thu Mar 14 00:01:31 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.552	0.562	-1.8	97	0.00
47 T	1,1,2-Trichloroethane	0.404	0.367	9.2	95	0.00
48 T	Dibromochloromethane	0.664	0.648	2.4	96	0.00
49 T	Bromoform	0.575	0.648	-12.7	97	0.00
50 T	4-Methyl-2-Pentanone	1.002	0.924	7.8	98	0.00
51 T	2-Hexanone	0.765	0.729	4.7	97	0.00
52 T	Tetrachloroethene	0.403	0.372	7.7	97	0.00
53 T	Toluene	1.125	1.062	5.6	95	0.00
54 T	1,2-Dibromoethane	0.644	0.587	8.9	96	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
56	1,1,1,2-Tetrachloroethane	0.449	0.410	8.7	96	0.00
57 T	Chlorobenzene	0.807	0.698	13.5	95	0.00
58 T	Ethyl Benzene	1.436	1.320	8.1	95	0.00
59 T	m/p-Xylene	1.207	1.099	8.9	95	0.00
60 T	o-Xylene	1.164	1.071	8.0	97	0.00
61 T	Styrene	0.817	0.800	2.1	95	0.00
62	Isopropylbenzene	1.638	1.510	7.8	97	0.00
63 T	1,1,2,2-Tetrachloroethane	0.839	0.729	13.1	96	0.00
64	n-propylbenzene	0.406	0.384	5.4	96	0.00
65	tert-Butylbenzene	1.384	1.330	3.9	97	0.00
66 T	Benzyl Chloride	0.551	0.677	-22.9	97	0.00
67	sec-Butylbenzene	1.990	1.903	4.4	98	0.00
68 S	1-Bromo-4-Fluorobenzene	0.764	0.709	7.2	98	0.00
69	p-Isopropyltoluene	1.590	1.565	1.6	98	0.00
70	n-Butylbenzene	1.663	1.648	0.9	98	0.00
71	2-Chlorotoluene	1.227	1.167	4.9	96	0.00
72 T	4-Ethyltoluene	1.247	1.247	0.0	97	0.00
73 T	1,3,5-Trimethylbenzene	1.173	1.197	-2.0	97	0.00
74 T	1,2,4-Trimethylbenzene	1.197	1.224	-2.3	98	0.00
75 T	1,3-Dichlorobenzene	0.749	0.692	7.6	97	0.00
76 T	1,4-Dichlorobenzene	0.731	0.696	4.8	98	0.00
77 T	1,2-Dichlorobenzene	0.729	0.680	6.7	98	0.00
78 T	Hexachloro-1,3-Butadiene	0.538	0.530	1.5	98	0.00
79 T	Naphthalene	0.983	1.269	-29.1	99	0.00
80 T	Naphthalene,2-methyl-	0.275	0.606	-120.4#	101	0.00
81 T	1,2,4-Trichlorobenzene	0.559	0.634	-13.4	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0