

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031424\
 Data File : VL041107.D
 Acq On : 14 Mar 2024 15:07
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/18/2024
 Supervised By :Mahesh Dadoda 03/18/2024

Quant Time: Mar 15 02:47:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 15 02:44:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1301356	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	3302933	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2850302	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	2217054	9.949	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%
Target Compounds						
20) Methylene Chloride	3.928	84	7428m	0.129	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031424\
Data File : VL041107.D
Acq On : 14 Mar 2024 15:07
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/18/2024
Supervised By :Mahesh Dadoda 03/18/2024

Quant Time: Mar 15 02:47:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 15 02:44:43 2024
Response via : Initial Calibration

