

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031424\
Data File : VL041114.D
Acq On : 14 Mar 2024 19:40
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/18/2024
Supervised By :Mahesh Dadoda 03/18/2024

Quant Time: Mar 15 02:50:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
QLast Update : Fri Mar 15 02:44:43 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	5.189	49	1254540	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	3226806	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2758183	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	2079719	9.644	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.400%
Target Compounds						
20) Methylene Chloride	3.931	84	7407m	0.133	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031424\
Data File : VL041114.D
Acq On : 14 Mar 2024 19:40
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/18/2024
Supervised By :Mahesh Dadoda 03/18/2024

Quant Time: Mar 15 02:50:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 15 02:44:43 2024
Response via : Initial Calibration

