

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031819\
 Data File : VL033362.D
 Acq On : 18 Mar 2019 17:27
 Operator : SY/AP
 Sample : 0.4 MDL
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 0.4 MDL

Manual Integrations
 APPROVED

MMDadoda
 3/21/2019 12:14:28 PM

Quant Time: Mar 19 07:13:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 1
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	781422	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1723314	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	1886893	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1452013	10.38	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds					Ovalue
13) Ethanol	3.66	45	13770	0.721 ppbv	99
40) 1,4-Dioxane	7.99	88	8658m	0.426 ppbv	
66) Benzyl Chloride	17.16	91	30650m	0.265 ppbv	
80) Naphthalene,2-methyl-	25.12	142	38961	0.325 ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031819\
Data File : VL033362.D
Acq On : 18 Mar 2019 17:27
Operator : SY/AP
Sample : 0.4 MDL
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
0.4 MDL

Manual Integrations
APPROVED

MMDadoda
3/21/2019 12:14:28 PM

Quant Time: Mar 19 07:13:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

