

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031819\
 Data File : VL033380.D
 Acq On : 19 Mar 2019 5:11
 Operator : SY/AP
 Sample : K1805-04DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ST-SS-1DL

Quant Time: Mar 19 10:22:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	812962	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1777087	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	1749097	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1332598	10.28	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds					Ovalue
52) Tetrachloroethene	11.37	164	380884	5.749 ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031819\
Data File : VL033380.D
Acq On : 19 Mar 2019 5:11
Operator : SY/AP
Sample : K1805-04DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ST-SS-1DL

Quant Time: Mar 19 10:22:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

