

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031819\
 Data File : VL033384.D
 Acq On : 19 Mar 2019 10:21
 Operator : SY/AP
 Sample : K1805-04 2X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ST-SS-1

Manual Integrations
 APPROVED

MMDadoda
 3/21/2019 12:15:59 PM

Quant Time: Mar 20 11:15:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 1
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	744894	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1655375	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.24	117	1766430	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1343956	10.27	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.08	85	27396	0.209	ppbv	95
3) Chlorodifluoromethane	3.01	51	11338m	0.133	ppbv	
11) Trichlorofluoromethane	4.00	101	14656	0.112	ppbv	94
13) Ethanol	3.65	45	13849m	0.761	ppbv	
15) Acetone	3.91	43	103621	0.895	ppbv	93
20) Methylene Chloride	4.40	84	19898m	0.527	ppbv	
26) Hexane	5.78	57	36409	0.484	ppbv #	38
38) Trichloroethene	7.95	130	54544	0.973	ppbv	90
52) Tetrachloroethene	11.38	164	7034883	113.983	ppbv	94
53) Toluene	9.94	91	114356	0.679	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031819\
Data File : VL033384.D
Acq On : 19 Mar 2019 10:21
Operator : SY/AP
Sample : K1805-04 2X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ST-SS-1

Manual Integrations
APPROVED

MMDadoda
3/21/2019 12:15:59 PM

Quant Time: Mar 20 11:15:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

