

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031921\
 Data File : VL036451.D
 Acq On : 19 Mar 2021 17:06
 Operator :
 Sample : OC030321 CAN 10448
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC030321 CAN 10448

Manual Integrations
 APPROVED

MMDadoda
 3/24/2021 6:14:43 PM

Quant Time: Mar 20 00:51:32 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar 0
 QLast Update : Wed Mar 03 14:18:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1716503	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.16	114	3872845	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.07	117	3454717	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2690696	9.47	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.70%

Target Compounds					Ovalue
15) Acetone	3.86	43	46860m	0.239	ppbv
20) Methylene Chloride	4.33	84	10971m	0.166	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031921\
Data File : VL036451.D
Acq On : 19 Mar 2021 17:06
Operator :
Sample : OC030321 CAN 10448
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC030321 CAN 10448

Manual Integrations
APPROVED

MMDadoda
3/24/2021 6:14:43 PM

Quant Time: Mar 20 00:51:32 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 03 14:18:24 2021
Response via : Initial Calibration

